


INTEGRIS PATIENT LIGHT 1  
BASIS HORIZONTAL ON WALL 1

SO13480330  
Unit ID:02A




Ceiling Line

PATIENT ROOMS

Please complete this room  
hanging chart.  
List the room #'s and check  
the desired unit handing.

3rd FLOOR...			4th FLOOR...		
Room #	As Shown	Opposite (Mirror Image)	Room #	As Shown	Opposite (Mirror Image)
329A			453A		
329B			453B		
330			454		
331			455		
371			428A		
372			428B		
327A			424		
327B			423		
326A			421		
326B			419		
325			418		
324			417		
322			416		
			415		
			414		



ITEM #	QTY	PART #	DESCRIPTION
LVC	1		Option
120	1		Voltage

ITEM #	QTY	PART #	DESCRIPTION
11	2		RECEPTACLE, DUPLEX IVORY (20A)
13	4		RECEPTACLE, DUPLEX RED (20A)
221	2		PROVISION, 1GANG W/FACEPLATE
224	1		NurseCall
558	2	P17070101	BOTTLE SLIDE STANDARD 1 3/16" INNER WIDTH
AIR	2		UNKNOWN Gas AIR UNKNOWN
O2	2		UNKNOWN Gas O2 UNKNOWN
VAC	2		UNKNOWN Gas VAC UNKNOWN

Supplemental Data:  
ITEM QTY PART# DESCRIPTION  
6 Basis GapTrim 0.345, Indent, Black  
Studs @ 3.625 Depth  
Metal Panels With: None 000  
P2102A201 Frame = 16x72x3.625  
Accessory Laminate Top Selection = Unspecified NONE 000  
Accessory Laminate Bottom Selection = Unspecified NONE 000  
MANIFOLD = P2102A000211

NOTE: NURSECALL LOCATED IN CENTER PANEL IS LIMITED  
TO A 3 GANG WITH A 2" DEEP BOX. NURSECALL  
LOCATED IN LEFT OR RIGHT PANEL LIMITED TO A 2 1/2"  
DEEP BOX.

ALL NURSE CALL PROVISIONS TO  
HAVE MOUNTING HOLES TAPPED BY  
OTHERS